



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/909,557	07/20/2001	Jean M. Beaupre	END-778	8228
27777	7590	10/28/2005	EXAMINER	
PHILIP S. JOHNSON JOHNSON & JOHNSON ONE JOHNSON & JOHNSON PLAZA NEW BRUNSWICK, NJ 08933-7003			DAVIS, DANIEL J	
			ART UNIT	PAPER NUMBER
			3733	
DATE MAILED: 10/28/2005				

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No.	Applicant(s)
	09/909,557	BEAUPRE, JEAN M.
	Examiner	Art Unit
	D. Jacob Davis	3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) Responsive to communication(s) filed on ____.
- 2a) This action is **FINAL**. 2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) Claim(s) 21-34 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) Claim(s) ____ is/are allowed.
- 6) Claim(s) 21-27 and 30-34 is/are rejected.
- 7) Claim(s) 28 and 29 is/are objected to.
- 8) Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on ____ is/are: a) accepted or b) objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) All b) Some * c) None of:
 1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. ____.
 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) Notice of References Cited (PTO-892)
- 2) Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
 Paper No(s)/Mail Date ____
- 4) Interview Summary (PTO-413)
 Paper No(s)/Mail Date ____
- 5) Notice of Informal Patent Application (PTO-152)
- 6) Other: ____

DETAILED ACTION

Claim Rejections - 35 USC § 102


The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 21-27 and 30-34 are rejected under 35 U.S.C. 102(b) as being anticipated by Idemoto (EP 0 45 470). Idemoto discloses a curved ultrasonic surgical end effector (Fig. 6B) comprising a concave treatment segment defining a first length and a central ridge contiguous with and extending substantially along the first length. (A figure illustrating the corresponding elements is found on page 3.) A plane located longitudinally along the treatment segment bisects the central ridge and the bottom surface of the device. The bottom is convex while the top is concave. Located on both said walls is a balance asymmetry. Because the balance asymmetries have more material toward the edges of the device and less in the middle, the asymmetries will counter torque.

Regarding claims 25-27, although Idemoto fails to disclose the balance ratio, the reference discloses the necessary features including balancing features, concavity, and a ridge. Therefore the device inherently has a balance ratio of less than 1:200.



Allowable Subject Matter

Claims 28 and 29 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims. The prior art fails to disclose or suggest all of the limitations of claim 28 including the asymmetry surface being substantially perpendicular to the plane of symmetry. The prior art fails to disclose or suggest all of the limitations of claim 29 including, "the first balance asymmetry being shorter than the second balance asymmetry."

Response to Arguments

Applicant's arguments filed November 15, 2004 have been fully considered but they are not persuasive. Examiner acknowledges applicant's traversal on the grounds that the Idemoto reference fails to disclose, "balance asymmetries" as required in the claims. Applicant does not recite the specifics of "a balance region," but generally recites "a balance region." Although Idemoto does not explicitly mention balance asymmetries as part of the device, the figure clearly illustrates asymmetrical portions on both sides of the end effector. Examiner has not used hindsight in the analysis since the reference anticipates the claimed device; there are no obviousness type rejections.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3733


the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to D. Jacob Davis whose telephone number is (571) 272-4693. The examiner can normally be reached on M-F.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo C. Robert can be reached on (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

DJD



EDUARDO C. ROBERT
PRIMARY EXAMINER